



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091923\
Data File : VY015635.D
Acq On : 19 Sep 2023 09:08
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Title : SW846 8260
Last Update : Thu Aug 31 20:01:47 2023



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	18442	PASS
75	95	30	60	53.4	50557	PASS
95	95	100	100	100.0	94701	PASS
96	95	5	9	6.6	6287	PASS
173	174	0.00	2	1.1	872	PASS
174	95	50	100	84.5	80069	PASS
175	174	5	9	7.6	6123	PASS
176	174	95	101	95.3	76328	PASS
177	176	5	9	7.2	5505	PASS