

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091924\
 Data File : VY019630.D
 Acq On : 19 Sep 2024 13:37
 Operator : SY/MD
 Sample : P4087-27
 Misc : 4.87g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-614-VOC-05

Quant Time: Sep 20 01:28:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	305081	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	586555	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	504101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	165745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	187361	66.323	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 132.640%	
35) Dibromofluoromethane	7.634	113	188297	56.619	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.240%	
50) Toluene-d8	10.109	98	690205	56.110	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.220%	
62) 4-Bromofluorobenzene	12.407	95	207432	43.857	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 87.720%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091924\
Data File : VY019630.D
Acq On : 19 Sep 2024 13:37
Operator : SY/MD
Sample : P4087-27
Misc : 4.87g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-614-VOC-05

Quant Time: Sep 20 01:28:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

