

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
 Data File : VY015656.D
 Acq On : 20 Sep 2023 14:08
 Operator : SY/MD
 Sample : 04436-17
 Misc : 4.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B102-10(8.5-9)

Quant Time: Sep 21 03:40:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	133570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	215002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	170020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	63831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	52786	35.918	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	71.840%
35) Dibromofluoromethane	7.716	113	62733	48.749	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.500%
50) Toluene-d8	10.179	98	248809	57.023	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	114.040%
62) 4-Bromofluorobenzene	12.483	95	64531	37.687	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.380%
Target Compounds						
17) Carbon Disulfide	4.192	76	91500	40.082	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
Data File : VY015656.D
Acq On : 20 Sep 2023 14:08
Operator : SY/MD
Sample : 04436-17
Misc : 4.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B102-10(8.5-9)

Quant Time: Sep 21 03:40:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

