

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
 Data File : VY015658.D
 Acq On : 20 Sep 2023 14:55
 Operator : SY/MD
 Sample : 04501-04
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Sep 21 03:41:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	124775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	213230	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	198299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	90933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	80415	58.574	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 117.140%	
35) Dibromofluoromethane	7.716	113	69780	54.676	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.360%	
50) Toluene-d8	10.179	98	251869	58.204	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 116.400%	
62) 4-Bromofluorobenzene	12.483	95	84552	49.790	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.580%	
Target Compounds						
20) Methylene Chloride	4.704	84	20895	9.019	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
Data File : VY015658.D
Acq On : 20 Sep 2023 14:55
Operator : SY/MD
Sample : 04501-04
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Sep 21 03:41:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

