

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
 Data File : VY015660.D
 Acq On : 20 Sep 2023 15:42
 Operator : SY/MD
 Sample : 04512-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11986

Quant Time: Sep 21 03:42:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	116176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	202610	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	185087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	85336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	76476	59.828	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 119.660%			
35) Dibromofluoromethane	7.716	113	64763	53.404	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 106.800%			
50) Toluene-d8	10.179	98	237692	57.807	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 115.620%			
62) 4-Bromofluorobenzene	12.477	95	79007	48.963	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 97.920%			
Target Compounds						
20) Methylene Chloride	4.704	84	25420	13.428	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
Data File : VY015660.D
Acq On : 20 Sep 2023 15:42
Operator : SY/MD
Sample : 04512-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11986

Quant Time: Sep 21 03:42:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

