

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
 Data File : VY015662.D
 Acq On : 20 Sep 2023 16:29
 Operator : SY/MD
 Sample : 04531-01
 Misc : 5.89g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-BCD

Quant Time: Sep 21 03:43:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	123965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	214618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	198975	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	93244	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	83282	61.059	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.120%
35) Dibromofluoromethane	7.716	113	69708	54.266	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.540%
50) Toluene-d8	10.179	98	255343	58.626	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	117.260%
62) 4-Bromofluorobenzene	12.477	95	86861	50.818	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092023\
Data File : VY015662.D
Acq On : 20 Sep 2023 16:29
Operator : SY/MD
Sample : 04531-01
Misc : 5.89g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-BCD

Quant Time: Sep 21 03:43:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

