

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
 Data File : VY006106.D
 Acq On : 22 Sep 2021 16:15
 Operator : SY/MD
 Sample : M3871-01
 Misc : 3.60G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS01-BL-(0.0)

Quant Time: Sep 23 01:27:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	150627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	237791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	209903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	86430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	102614	49.044	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.080%
35) Dibromofluoromethane	7.728	113	73862	50.009	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.020%
50) Toluene-d8	10.185	98	295603	46.959	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.920%
62) 4-Bromofluorobenzene	12.483	95	88014	41.182	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.360%
Target Compounds						Qvalue
16) Acetone	3.960	43	55853	97.326	ug/l	100
17) Carbon Disulfide	4.204	76	7735	1.540	ug/l	99
25) 2-Butanone	6.996	43	9504	12.016	ug/l	98
86) p-Isopropyltoluene	13.373	119	9282	1.513	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
Data File : VY006106.D
Acq On : 22 Sep 2021 16:15
Operator : SY/MD
Sample : M3871-01
Misc : 3.60G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS01-BL-(0.0)

Quant Time: Sep 23 01:27:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

