

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
 Data File : VY006108.D
 Acq On : 22 Sep 2021 17:01
 Operator : SY/MD
 Sample : M3871-03
 Misc : 4.91G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS03-BL-(0.0)

Quant Time: Sep 23 01:27:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	151522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	245862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	218833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	91340	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	102831	48.857	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.720%
35) Dibromofluoromethane	7.728	113	73705	48.265	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.520%
50) Toluene-d8	10.185	98	308315	47.371	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.740%
62) 4-Bromofluorobenzene	12.483	95	92660	41.932	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
Data File : VY006108.D
Acq On : 22 Sep 2021 17:01
Operator : SY/MD
Sample : M3871-03
Misc : 4.91G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS03-BL-(0.0)

Quant Time: Sep 23 01:27:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

