

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
 Data File : VY006102.D
 Acq On : 22 Sep 2021 13:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 22 18:13:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.528	0.470	11.0	104	0.00
3 P	Chloromethane	0.657	0.562	14.5	100	0.00
4 C	Vinyl Chloride	0.742	0.658	11.3#	102	0.00
5 T	Bromomethane	0.481	0.426	11.4	105	0.00
6 T	Chloroethane	0.445	0.398	10.6	101	0.00
7 T	Trichlorofluoromethane	0.967	0.884	8.6	103	0.00
8 T	Diethyl Ether	0.318	0.264	17.0	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.472	2.3	111	0.00
10 T	Methyl Iodide	0.511	0.484	5.3	102	0.00
11 T	Tert butyl alcohol	0.142	0.112	21.1	124	-0.01
12 CM	1,1-Dichloroethene	0.499	0.477	4.4#	109	0.00
13 T	Acrolein	0.042	0.047	-11.9	120	0.00
14 T	Allyl chloride	1.066	0.969	9.1	102	0.00
15 T	Acrylonitrile	0.181	0.180	0.6	103	0.00
16 T	Acetone	0.190	0.213	-12.1	116	0.00
17 T	Carbon Disulfide	1.667	1.561	6.4	105	0.00
18 T	Methyl Acetate	0.480	0.447	6.9	100	0.00
19 T	Methyl tert-butyl Ether	1.609	1.295	19.5	86	0.00
20 T	Methylene Chloride	0.707	0.628	11.2	117	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.524	6.4	104	0.00
22 T	Diisopropyl ether	2.071	1.677	19.0	88	0.00
23 T	Vinyl Acetate	1.307	1.113	14.8	89	0.00
24 P	1,1-Dichloroethane	1.061	0.960	9.5	101	0.00
25 T	2-Butanone	0.263	0.265	-0.8	104	0.00
26 T	2,2-Dichloropropane	1.005	0.932	7.3	106	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.571	8.6	100	0.00
28 T	Bromochloromethane	0.466	0.434	6.9	94	0.00
29 T	Tetrahydrofuran	0.165	0.159	3.6	98	0.00
30 C	Chloroform	1.064	0.935	12.1#	98	0.00
31 T	Cyclohexane	1.113	0.980	11.9	103	0.00
32 T	1,1,1-Trichloroethane	0.999	0.893	10.6	99	0.00
33 S	1,2-Dichloroethane-d4	0.695	0.558	19.7	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.311	0.287	7.7	99	0.00
36 T	1,1-Dichloropropene	0.540	0.525	2.8	103	0.00
37 T	Ethyl Acetate	0.373	0.358	4.0	98	0.00
38 T	Carbon Tetrachloride	0.588	0.552	6.1	100	0.00
39 T	Methylcyclohexane	0.669	0.652	2.5	103	0.00
40 TM	Benzene	1.548	1.454	6.1	100	0.00
41 T	Methacrylonitrile	0.218	0.228	-4.6	100	0.01
42 TM	1,2-Dichloroethane	0.547	0.456	16.6	86	0.00
43 T	Isopropyl Acetate	0.694	0.618	11.0	89	0.00
44 TM	Trichloroethene	0.386	0.381	1.3	104	0.00
45 C	1,2-Dichloropropane	0.383	0.346	9.7#	96	0.00
46 T	Dibromomethane	0.207	0.185	10.6	91	0.00
47 T	Bromodichloromethane	0.533	0.472	11.4	92	0.00
48 T	Methyl methacrylate	0.321	0.297	7.5	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
 Data File : VY006102.D
 Acq On : 22 Sep 2021 13:36
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 22 18:13:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	109	0.00
50 S	Toluene-d8	1.324	1.240	6.3	99	0.00
51 T	4-Methyl-2-Pentanone	0.370	0.344	7.0	92	0.00
52 CM	Toluene	0.954	0.889	6.8#	99	0.00
53 T	t-1,3-Dichloropropene	0.586	0.514	12.3	90	0.00
54 T	cis-1,3-Dichloropropene	0.632	0.567	10.3	94	0.00
55 T	1,1,2-Trichloroethane	0.285	0.256	10.2	91	0.00
56 T	Ethyl methacrylate	0.455	0.393	13.6	86	0.00
57 T	1,3-Dichloropropane	0.525	0.463	11.8	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.238	0.216	9.2	93	0.00
59 T	2-Hexanone	0.256	0.258	-0.8	97	0.00
60 T	Dibromochloromethane	0.351	0.314	10.5	92	0.00
61 T	1,2-Dibromoethane	0.269	0.247	8.2	93	0.00
62 S	4-Bromofluorobenzene	0.449	0.400	10.9	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.409	0.421	-2.9	105	0.00
65 PM	Chlorobenzene	1.070	1.011	5.5	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.399	0.363	9.0	92	0.00
67 C	Ethyl Benzene	2.034	1.983	2.5#	98	0.00
68 T	m/p-Xylenes	0.771	0.750	2.7	98	0.00
69 T	o-Xylene	0.728	0.695	4.5	95	0.00
70 T	Styrene	1.248	1.187	4.9	94	0.00
71 P	Bromoform	0.250	0.243	2.8	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.915	3.859	1.4	98	0.00
74 T	N-amyl acetate	1.352	1.186	12.3	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.728	0.703	3.4	93	0.00
76 T	1,2,3-Trichloropropane	0.533	0.637	-19.5	122	0.00
77 T	Bromobenzene	0.914	0.853	6.7	92	0.00
78 T	n-propylbenzene	4.844	4.781	1.3	98	0.00
79 T	2-Chlorotoluene	2.634	2.602	1.2	97	0.00
80 T	1,3,5-Trimethylbenzene	3.247	3.181	2.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.290	0.291	-0.3	95	0.00
82 T	4-Chlorotoluene	2.723	2.689	1.2	98	0.00
83 T	tert-Butylbenzene	2.777	2.779	-0.1	97	0.00
84 T	1,2,4-Trimethylbenzene	3.222	3.128	2.9	94	0.00
85 T	sec-Butylbenzene	4.158	4.206	-1.2	98	0.00
86 T	p-Isopropyltoluene	3.549	3.481	1.9	96	0.00
87 T	1,3-Dichlorobenzene	1.762	1.704	3.3	94	0.00
88 T	1,4-Dichlorobenzene	1.767	1.677	5.1	94	0.00
89 T	n-Butylbenzene	3.319	3.370	-1.5	99	0.00
90 T	Hexachloroethane	0.639	0.636	0.5	98	0.00
91 T	1,2-Dichlorobenzene	1.547	1.463	5.4	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.145	0.145	0.0	96	0.00
93 T	1,2,4-Trichlorobenzene	1.004	0.965	3.9	94	0.00
94 T	Hexachlorobutadiene	0.650	0.658	-1.2	101	0.00
95 T	Naphthalene	1.998	2.051	-2.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092221\
Data File : VY006102.D
Acq On : 22 Sep 2021 13:36
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 22 18:13:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.870	0.815	6.3	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6