

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092223\
 Data File : VY015687.D
 Acq On : 22 Sep 2023 16:17
 Operator : SY/MD
 Sample : 04582-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-092223

Quant Time: Sep 22 22:57:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	130804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	225167	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	208146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	95407	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	87228	60.608	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.220%
35) Dibromofluoromethane	7.716	113	73506	54.542	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.080%
50) Toluene-d8	10.179	98	265777	58.162	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.320%
62) 4-Bromofluorobenzene	12.483	95	88456	49.327	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092223\
Data File : VY015687.D
Acq On : 22 Sep 2023 16:17
Operator : SY/MD
Sample : 04582-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-092223

Quant Time: Sep 22 22:57:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

