

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092321\
 Data File : VY006116.D
 Acq On : 23 Sep 2021 11:38
 Operator : SY/MD
 Sample : VY0923SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0923SBL01

Quant Time: Sep 24 07:26:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	122945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	196074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	182706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84644	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	79222	46.389	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.780%
35) Dibromofluoromethane	7.722	113	56742	46.592	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.180%
50) Toluene-d8	10.179	98	242056	46.634	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.260%
62) 4-Bromofluorobenzene	12.483	95	77314	43.872	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092321\
Data File : VY006116.D
Acq On : 23 Sep 2021 11:38
Operator : SY/MD
Sample : VY0923SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0923SBL01

Quant Time: Sep 24 07:26:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

