

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092321\
 Data File : VY006123.D
 Acq On : 23 Sep 2021 14:47
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Sep 24 07:29:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	157689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	247008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	233153	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	117127	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	98362	44.906	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.820%
35) Dibromofluoromethane	7.728	113	73094	47.643	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.280%
50) Toluene-d8	10.185	98	319372	48.842	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.680%
62) 4-Bromofluorobenzene	12.483	95	111323	50.144	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092321\
Data File : VY006123.D
Acq On : 23 Sep 2021 14:47
Operator : SY/MD
Sample : VIBLK
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Sep 24 07:29:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 980
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 157689
Ion Ratio Lower Upper
168 100
99 54.3 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 44.906 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47

Tgt Ion: 65 Resp: 98362
Ion Ratio Lower Upper
65 100
67 52.5 0.0 99.4









#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK




Tgt Ion:	114	Resp:	247008
Ion	Ratio	Lower	Upper
114	100		
63	21.0	0.0	43.2
88	15.4	0.0	31.8




#35

Dibromofluoromethane
Concen: 47.643 ug/l
RT: 7.728 min Scan# 969
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47

Tgt Ion:	113	Resp:	73094
Ion	Ratio	Lower	Upper
113	100		
111	102.4	82.6	124.0
192	22.4	17.6	26.4









#50
Toluene-d8
Concen: 48.842 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47




Instrument :
MSVOA_Y
ClientSampled :
VIBLK


Tgt Ion: 98 Resp: 319372
Ion Ratio Lower Upper
98 100
100 62.5 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 50.144 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47

Tgt Ion: 95 Resp: 111323
Ion Ratio Lower Upper
95 100
174 87.0 0.0 177.0
176 84.3 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Tgt Ion:	117	Resp:	233153
Ion	Ratio	Lower	Upper
117	100		
82	54.5	43.0	64.4
119	31.2	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY006123.D
Acq: 23 Sep 2021 14:47

Tgt Ion:	152	Resp:	117127
Ion	Ratio	Lower	Upper
152	100		
115	56.5	28.2	84.7
150	155.8	0.0	347.6

