

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092421\
 Data File : VY006139.D
 Acq On : 24 Sep 2021 15:32
 Operator : SY/MD
 Sample : M3917-07
 Misc : 5.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-10

Quant Time: Sep 25 05:45:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	47059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	62881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	51347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	22536	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	26366	40.335	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.660%
35) Dibromofluoromethane	7.722	113	20319	52.025	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.040%
50) Toluene-d8	10.185	98	76101	45.717	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.440%
62) 4-Bromofluorobenzene	12.483	95	20316	35.947	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	71.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092421\
Data File : VY006139.D
Acq On : 24 Sep 2021 15:32
Operator : SY/MD
Sample : M3917-07
Misc : 5.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-10

Quant Time: Sep 25 05:45:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

