

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092421\
 Data File : VY006150.D
 Acq On : 24 Sep 2021 19:48
 Operator : SY/MD
 Sample : M3934-01
 Misc : 5.01G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-5

Quant Time: Sep 25 05:48:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	6465	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	10034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	9277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5389	60.009	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.020%	
35) Dibromofluoromethane	7.728	113	3181	51.041	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.080%	
50) Toluene-d8	10.185	98	12221	46.009	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.020%	
62) 4-Bromofluorobenzene	12.483	95	3613	40.063	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 80.120%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	874	35.484	ug/l #	80
20) Methylene Chloride	4.710	84	919	8.059	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092421\
Data File : VY006150.D
Acq On : 24 Sep 2021 19:48
Operator : SY/MD
Sample : M3934-01
Misc : 5.01G/5ML/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-5

Quant Time: Sep 25 05:48:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

