

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092421\
 Data File : VY006134.D
 Acq On : 24 Sep 2021 13:36
 Operator : SY/MD
 Sample : M3916-07
 Misc : 5.06G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CONC-3

Quant Time: Sep 25 05:44:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	119719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	187871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	170117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	71440	42.959	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.920%
35) Dibromofluoromethane	7.722	113	30207	25.887	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	51.780%#
50) Toluene-d8	10.185	98	237583	47.771	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.540%
62) 4-Bromofluorobenzene	12.483	95	71736	42.484	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.960%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092421\
Data File : VY006134.D
Acq On : 24 Sep 2021 13:36
Operator : SY/MD
Sample : M3916-07
Misc : 5.06G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CONC-3

Quant Time: Sep 25 05:44:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.000 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36




Instrument :
MSVOA_Y
ClientSampleId :
CONC-3


Tgt Ion:168 Resp: 119719
Ion Ratio Lower Upper
168 100
99 56.6 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 42.959 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. 0.000 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36

Tgt Ion: 65 Resp: 71440
Ion Ratio Lower Upper
65 100
67 51.5 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36




Instrument :
MSVOA_Y
ClientSampled :
CONC-3


Tgt Ion:	114	Resp:	187871
Ion	Ratio	Lower	Upper
114	100		
63	22.1	0.0	43.2
88	15.0	0.0	31.8



#35
Dibromofluoromethane
Concen: 25.887 ug/l
RT: 7.722 min Scan# 968
Delta R.T. -0.006 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36

Tgt Ion:	113	Resp:	30207
Ion	Ratio	Lower	Upper
113	100		
111	104.8	82.6	124.0
192	22.6	17.6	26.4









#50
Toluene-d8
Concen: 47.771 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36




Instrument :
MSVOA_Y
ClientSampleId :
CONC-3


Tgt Ion: 98 Resp: 237583
Ion Ratio Lower Upper
98 100
100 63.3 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 42.484 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36

Tgt Ion: 95 Resp: 71736
Ion Ratio Lower Upper
95 100
174 91.4 0.0 177.0
176 88.6 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. -0.006 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36

Instrument :
MSVOA_Y
ClientSampled :
CONC-3

Tgt Ion:	117	Resp:	170117
Ion	Ratio	Lower	Upper
117	100		
82	57.5	43.0	64.4
119	32.1	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY006134.D
Acq: 24 Sep 2021 13:36

Tgt Ion:	152	Resp:	77436
Ion	Ratio	Lower	Upper
152	100		
115	56.1	28.2	84.7
150	158.3	0.0	347.6

