

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092523\
 Data File : VY015696.D
 Acq On : 25 Sep 2023 10:51
 Operator : SY/MD
 Sample : VY0925SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0925SBL01

Quant Time: Sep 26 04:22:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	129314	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	220667	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.495	117	201589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	92653	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	82372	57.894	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.780%
35) Dibromofluoromethane	7.722	113	72136	54.617	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.240%
50) Toluene-d8	10.185	98	261761	58.452	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.900%
62) 4-Bromofluorobenzene	12.489	95	85389	48.588	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092523\
Data File : VY015696.D
Acq On : 25 Sep 2023 10:51
Operator : SY/MD
Sample : VY0925SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0925SBL01

Quant Time: Sep 26 04:22:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

