

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092523\
 Data File : VY015700.D
 Acq On : 25 Sep 2023 13:33
 Operator : SY/MD
 Sample : 04566-03
 Misc : 4.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CPT102-06(8.5-9)

Quant Time: Sep 26 04:23:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	124846	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.691	114	215021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	193983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	83263	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	81658	59.446	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.900%
35) Dibromofluoromethane	7.716	113	69662	54.128	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.260%
50) Toluene-d8	10.185	98	254167	58.246	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.500%
62) 4-Bromofluorobenzene	12.489	95	80668	47.107	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.220%
Target Compounds						
16) Acetone	3.966	43	8986	12.224	ug/l	97
17) Carbon Disulfide	4.198	76	30098	14.106	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092523\
Data File : VY015700.D
Acq On : 25 Sep 2023 13:33
Operator : SY/MD
Sample : 04566-03
Misc : 4.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CPT102-06(8.5-9)

Quant Time: Sep 26 04:23:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

