

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
 Data File : VY015735.D
 Acq On : 27 Sep 2023 19:28
 Operator : SY/MD
 Sample : 04627-19
 Misc : 5.75g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 280-300FB-BACKFILL-10

Quant Time: Sep 28 00:49:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	122283	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.691	114	215987	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	201633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92700	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	83815	62.295	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.580%
35) Dibromofluoromethane	7.722	113	69246	53.565	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.120%
50) Toluene-d8	10.185	98	253753	57.891	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.780%
62) 4-Bromofluorobenzene	12.483	95	85101	49.473	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
Data File : VY015735.D
Acq On : 27 Sep 2023 19:28
Operator : SY/MD
Sample : 04627-19
Misc : 5.75g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
280-300FB-BACKFILL-10

Quant Time: Sep 28 00:49:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

