

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
 Data File : VY015737.D
 Acq On : 27 Sep 2023 20:15
 Operator : SY/MD
 Sample : 04627-23
 Misc : 5.48g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 280-300FB-BACKFILL-12

Quant Time: Sep 28 00:50:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	116984	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	206983	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.495	117	192776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87701	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	80290	62.378	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.760%
35) Dibromofluoromethane	7.728	113	68097	54.967	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.940%
50) Toluene-d8	10.185	98	244750	58.266	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.540%
62) 4-Bromofluorobenzene	12.489	95	80486	48.826	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
Data File : VY015737.D
Acq On : 27 Sep 2023 20:15
Operator : SY/MD
Sample : 04627-23
Misc : 5.48g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
280-300FB-BACKFILL-12

Quant Time: Sep 28 00:50:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

