

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
 Data File : VY015720.D
 Acq On : 27 Sep 2023 09:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 28 00:43:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.01
2 T	Dichlorodifluoromethane	0.330	0.325	1.5	90	0.01
3 P	Chloromethane	0.416	0.401	3.6	89	0.00
4 C	Vinyl Chloride	0.475	0.489	-2.9#	94	0.01
5 T	Bromomethane	0.361	0.362	-0.3	89	0.02
6 T	Chloroethane	0.352	0.351	0.3	92	0.01
7 T	Trichlorofluoromethane	0.797	0.875	-9.8	105	0.01
8 T	Diethyl Ether	0.340	0.329	3.2	92	0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.499	0.509	-2.0	97	0.01
10 T	Methyl Iodide	0.519	0.527	-1.5	86	0.01
11 T	Tert butyl alcohol	0.174	0.114	34.5#	106	0.05
12 CM	1,1-Dichloroethene	0.440	0.443	-0.7#	96	0.01
13 T	Acrolein	0.084	0.077	8.3	90	0.02
14 T	Allyl chloride	0.804	0.779	3.1	91	0.01
15 T	Acrylonitrile	0.249	0.230	7.6	92	0.02
16 T	Acetone	0.294	0.331	-12.6	116	0.03
17 T	Carbon Disulfide	0.855	0.822	3.9	86	0.01
18 T	Methyl Acetate	0.988	0.932	5.7	92	0.02
19 T	Methyl tert-butyl Ether	1.792	1.764	1.6	94	0.02
20 T	Methylene Chloride	0.779	0.664	14.8	101	0.02
21 T	trans-1,2-Dichloroethene	0.503	0.499	0.8	91	0.02
22 T	Diisopropyl ether	1.936	1.899	1.9	91	0.02
23 T	Vinyl Acetate	1.195	1.137	4.9	90	0.02
24 P	1,1-Dichloroethane	1.055	1.036	1.8	93	0.02
25 T	2-Butanone	0.416	0.406	2.4	103	0.02
26 T	2,2-Dichloropropane	0.964	0.949	1.6	101	0.02
27 T	cis-1,2-Dichloroethene	0.667	0.656	1.6	93	0.02
28 T	Bromochloromethane	0.286	0.268	6.3	91	0.02
29 T	Tetrahydrofuran	0.235	0.213	9.4	90	0.02
30 C	Chloroform	1.130	1.112	1.6#	95	0.02
31 T	Cyclohexane	0.762	0.711	6.7	92	0.01
32 T	1,1,1-Trichloroethane	0.933	0.958	-2.7	98	0.02
33 S	1,2-Dichloroethane-d4	0.550	0.563	-2.4	112	0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.01
35 S	Dibromofluoromethane	0.299	0.322	-7.7	110	0.01
36 T	1,1-Dichloropropene	0.442	0.454	-2.7	95	0.01
37 T	Ethyl Acetate	0.441	0.441	0.0	97	0.02
38 T	Carbon Tetrachloride	0.480	0.507	-5.6	98	0.02
39 T	Methylcyclohexane	0.489	0.502	-2.7	94	0.01
40 TM	Benzene	1.349	1.365	-1.2	93	0.02
41 T	Methacrylonitrile	0.238	0.207	13.0	94	0.00
42 TM	1,2-Dichloroethane	0.462	0.459	0.6	93	0.01
43 T	Isopropyl Acetate	0.754	0.743	1.5	94	0.02
44 TM	Trichloroethene	0.377	0.385	-2.1	94	0.01
45 C	1,2-Dichloropropane	0.376	0.380	-1.1#	92	0.00
46 T	Dibromomethane	0.241	0.240	0.4	92	0.02
47 T	Bromodichloromethane	0.537	0.550	-2.4	95	0.01
48 T	Methyl methacrylate	0.331	0.324	2.1	92	0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
 Data File : VY015720.D
 Acq On : 27 Sep 2023 09:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 28 00:43:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.005	0.0	100	0.04
50 S	Toluene-d8	1.015	1.148	-13.1	122	0.00
51 T	4-Methyl-2-Pentanone	0.467	0.458	1.9	93	0.00
52 CM	Toluene	0.870	0.883	-1.5#	92	0.01
53 T	t-1,3-Dichloropropene	0.575	0.583	-1.4	93	0.01
54 T	cis-1,3-Dichloropropene	0.619	0.623	-0.6	93	0.01
55 T	1,1,2-Trichloroethane	0.353	0.351	0.6	93	0.00
56 T	Ethyl methacrylate	0.520	0.531	-2.1	93	0.00
57 T	1,3-Dichloropropane	0.592	0.592	0.0	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.165	0.156	5.5	94	0.01
59 T	2-Hexanone	0.359	0.373	-3.9	100	0.00
60 T	Dibromochloromethane	0.412	0.419	-1.7	94	0.00
61 T	1,2-Dibromoethane	0.345	0.338	2.0	93	0.00
62 S	4-Bromofluorobenzene	0.398	0.425	-6.8	112	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.429	0.459	-7.0	97	0.01
65 PM	Chlorobenzene	1.091	1.109	-1.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.424	0.442	-4.2	94	0.00
67 C	Ethyl Benzene	1.908	1.986	-4.1#	93	0.01
68 T	m/p-Xylenes	0.725	0.755	-4.1	93	0.00
69 T	o-Xylene	0.711	0.748	-5.2	94	0.00
70 T	Styrene	1.217	1.287	-5.8	93	0.00
71 P	Bromoform	0.307	0.326	-6.2	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.751	3.896	-3.9	95	0.01
74 T	N-amyl acetate	1.410	1.440	-2.1	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.044	1.041	0.3	95	0.00
76 T	1,2,3-Trichloropropane	0.798	0.800	-0.3	91	0.00
77 T	Bromobenzene	0.921	0.938	-1.8	94	0.00
78 T	n-propylbenzene	4.557	4.721	-3.6	93	0.00
79 T	2-Chlorotoluene	2.597	2.654	-2.2	92	0.01
80 T	1,3,5-Trimethylbenzene	3.140	3.257	-3.7	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.354	0.365	-3.1	94	0.00
82 T	4-Chlorotoluene	2.712	2.776	-2.4	92	0.01
83 T	tert-Butylbenzene	2.750	2.916	-6.0	96	0.00
84 T	1,2,4-Trimethylbenzene	3.142	3.254	-3.6	92	0.00
85 T	sec-Butylbenzene	4.116	4.310	-4.7	94	0.00
86 T	p-Isopropyltoluene	3.438	3.586	-4.3	93	0.00
87 T	1,3-Dichlorobenzene	1.827	1.865	-2.1	93	0.01
88 T	1,4-Dichlorobenzene	1.827	1.850	-1.3	92	0.00
89 T	n-Butylbenzene	3.256	3.356	-3.1	92	0.00
90 T	Hexachloroethane	0.664	0.698	-5.1	95	0.00
91 T	1,2-Dichlorobenzene	1.686	1.733	-2.8	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.207	0.205	1.0	96	0.00
93 T	1,2,4-Trichlorobenzene	1.062	1.087	-2.4	92	0.00
94 T	Hexachlorobutadiene	0.544	0.532	2.2	87	0.00
95 T	Naphthalene	2.768	2.845	-2.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092723\
Data File : VY015720.D
Acq On : 27 Sep 2023 09:32
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 28 00:43:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.980	0.992	-1.2	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6