

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092821\
 Data File : VY006199.D
 Acq On : 28 Sep 2021 19:28
 Operator : SY/MD
 Sample : M3971-10
 Misc : 4.63G/5ML/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 KEARNY-COMP

Quant Time: Sep 29 03:22:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	8742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	13484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	11797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4665	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5808	62.075	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.140%
35) Dibromofluoromethane	7.728	113	2827	39.242	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.480%
50) Toluene-d8	10.185	98	16402	54.401	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.800%
62) 4-Bromofluorobenzene	12.483	95	4655	45.449	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092821\
Data File : VY006199.D
Acq On : 28 Sep 2021 19:28
Operator : SY/MD
Sample : M3971-10
Misc : 4.63G/5ML/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
KEARNY-COMP

Quant Time: Sep 29 03:22:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

