

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092823\
 Data File : VY015751.D
 Acq On : 28 Sep 2023 18:18
 Operator : SY/MD
 Sample : 04627-17RE
 Misc : 3.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 280-300FB-BACKFILL-09RE

Quant Time: Sep 29 01:27:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	113875	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.697	114	203234	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.495	117	191845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	87322	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	80000	63.850	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.700%
35) Dibromofluoromethane	7.728	113	68148	56.023	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.040%
50) Toluene-d8	10.185	98	243554	59.051	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	118.100%
62) 4-Bromofluorobenzene	12.483	95	80346	49.640	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092823\
Data File : VY015751.D
Acq On : 28 Sep 2023 18:18
Operator : SY/MD
Sample : 04627-17RE
Misc : 3.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
280-300FB-BACKFILL-09RE

Quant Time: Sep 29 01:27:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

