

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
 Data File : VY006205.D
 Acq On : 29 Sep 2021 12:15
 Operator : SY/MD
 Sample : VY0929SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0929SBL01

Quant Time: Sep 30 06:59:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	177852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	282856	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	255839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	124077	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	92931	48.820	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.640%
35) Dibromofluoromethane	7.722	113	72502	47.977	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.960%
50) Toluene-d8	10.185	98	302359	47.807	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.483	95	99002	46.079	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
Data File : VY006205.D
Acq On : 29 Sep 2021 12:15
Operator : SY/MD
Sample : VY0929SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0929SBL01

Quant Time: Sep 30 06:59:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

