

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
 Data File : VY006209.D
 Acq On : 29 Sep 2021 13:46
 Operator : SY/MD
 Sample : M3971-01RE
 Misc : 2.10G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-COMP-1RE

Quant Time: Sep 30 07:00:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	82154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	133851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	123719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	56501	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	53198	60.501	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.000%
35) Dibromofluoromethane	7.728	113	40124	56.109	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.220%
50) Toluene-d8	10.185	98	165819	55.405	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.800%
62) 4-Bromofluorobenzene	12.483	95	54176	53.285	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
Data File : VY006209.D
Acq On : 29 Sep 2021 13:46
Operator : SY/MD
Sample : M3971-01RE
Misc : 2.10G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-COMP-1RE

Quant Time: Sep 30 07:00:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

