

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
 Data File : VY006216.D
 Acq On : 29 Sep 2021 16:28
 Operator : SY/MD
 Sample : M3987-02
 Misc : 3.79G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS03-NA

Quant Time: Sep 30 07:01:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	149039	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	247423	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	228159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	102003	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	96632	60.58	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.16%	
35) Dibromofluoromethane	7.728	113	74684	56.50	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.00%	
50) Toluene-d8	10.185	98	312166	56.43	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.86%	
62) 4-Bromofluorobenzene	12.483	95	97124	51.68	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.36%	
Target Compounds						
16) Acetone	3.979	43	13508	30.32	ug/l	99
17) Carbon Disulfide	4.204	76	5206	1.12	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
Data File : VY006216.D
Acq On : 29 Sep 2021 16:28
Operator : SY/MD
Sample : M3987-02
Misc : 3.79G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS03-NA

Quant Time: Sep 30 07:01:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

