

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
 Data File : VY006220.D
 Acq On : 29 Sep 2021 18:01
 Operator : SY/MD
 Sample : M3966-03
 Misc : 6.17G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-830-J-SO-1-1.5-092721

Quant Time: Sep 30 07:34:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	146963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	244152	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	218454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	86499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	102509	65.171	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.340%
35) Dibromofluoromethane	7.722	113	73009	55.971	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.940%
50) Toluene-d8	10.185	98	303809	55.651	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.300%
62) 4-Bromofluorobenzene	12.483	95	89270	48.135	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.280%
Target Compounds						
					Qvalue	
16) Acetone	3.967	43	1921	4.373	ug/l	94
44) Trichloroethene	8.947	130	8196	4.605	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
Data File : VY006220.D
Acq On : 29 Sep 2021 18:01
Operator : SY/MD
Sample : M3966-03
Misc : 6.17G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-830-J-SO-1-1.5-092721

Quant Time: Sep 30 07:34:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

