

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
 Data File : VY006227.D
 Acq On : 29 Sep 2021 20:44
 Operator : SY/MD
 Sample : M3966-01
 Misc : 7.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-831-J-SO-1-1.5-092721

Quant Time: Sep 30 07:35:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	140249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	230996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	213375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	94329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	97564	64.996	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.000%	
35) Dibromofluoromethane	7.728	113	69576	56.377	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.760%	
50) Toluene-d8	10.185	98	289674	56.084	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.160%	
62) 4-Bromofluorobenzene	12.483	95	91303	52.036	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.080%	
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	6.997	96	1649	1.011	ug/l	75
44) Trichloroethene	8.941	130	4520	2.684	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
Data File : VY006227.D
Acq On : 29 Sep 2021 20:44
Operator : SY/MD
Sample : M3966-01
Misc : 7.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-831-J-SO-1-1.5-092721

Quant Time: Sep 30 07:35:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

