



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092921\
Data File : VY006203.D
Acq On : 29 Sep 2021 10:30
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y0928215.M
Title : SW846 8260
Last Update : Tue Sep 28 17:44:36 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.8	18160	PASS
75	95	30	60	55.0	45848	PASS
95	95	100	100	100.0	83299	PASS
96	95	5	9	6.6	5528	PASS
173	174	0.00	2	0.4	261	PASS
174	95	50	100	84.2	70099	PASS
175	174	5	9	8.1	5711	PASS
176	174	95	101	95.3	66808	PASS
177	176	5	9	6.7	4509	PASS