



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092923\
Data File : VY015755.D
Acq On : 29 Sep 2023 12:12
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Title : SW846 8260
Last Update : Wed Sep 20 01:39:40 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	19981	PASS
75	95	30	60	53.9	55123	PASS
95	95	100	100	100.0	102189	PASS
96	95	5	9	6.8	6900	PASS
173	174	0.00	2	0.8	682	PASS
174	95	50	100	81.6	83360	PASS
175	174	5	9	7.6	6341	PASS
176	174	95	101	95.9	79971	PASS
177	176	5	9	6.8	5472	PASS