

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093021\
 Data File : VY006236.D
 Acq On : 30 Sep 2021 13:46
 Operator : SY/MD
 Sample : M3990-03
 Misc : 7.24G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-840-R-SO-6-6.5-092921

Quant Time: Oct 01 06:27:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	129665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	211086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	197841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	85533	61.632	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 123.260%		
35) Dibromofluoromethane	7.728	113	62766	55.656	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 111.320%		
50) Toluene-d8	10.185	98	262793	55.678	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 111.360%		
62) 4-Bromofluorobenzene	12.483	95	84222	52.527	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 105.060%		
Target Compounds						
16) Acetone	3.991	43	6429	16.586	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093021\
Data File : VY006236.D
Acq On : 30 Sep 2021 13:46
Operator : SY/MD
Sample : M3990-03
Misc : 7.24G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-840-R-SO-6-6.5-092921

Quant Time: Oct 01 06:27:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

