

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093021\
 Data File : VY006246.D
 Acq On : 30 Sep 2021 17:38
 Operator : SY/MD
 Sample : M3998-04
 Misc : 5.01G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MIXED-DEBRIS-2

Quant Time: Oct 01 06:28:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	41966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	67464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	65069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	29872	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	30354	67.580	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.160%
35) Dibromofluoromethane	7.728	113	20918	58.036	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.080%
50) Toluene-d8	10.185	98	84686	56.140	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.280%
62) 4-Bromofluorobenzene	12.483	95	26295	51.312	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093021\
Data File : VY006246.D
Acq On : 30 Sep 2021 17:38
Operator : SY/MD
Sample : M3998-04
Misc : 5.01G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MIXED-DEBRIS-2

Quant Time: Oct 01 06:28:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

