

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
 Data File : VY010753.D
 Acq On : 30 Sep 2022 16:38
 Operator : KP/MD
 Sample : N4898-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CRUSHED-MASONRY-PILE-11RE

Quant Time: Oct 01 06:19:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	220441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	367100	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	327010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	148805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	119516	56.684	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.360%
35) Dibromofluoromethane	7.709	113	5781	2.468	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	4.940%#
50) Toluene-d8	10.179	98	396418	46.046	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.100%
62) 4-Bromofluorobenzene	12.477	95	140812	43.467	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.940%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
Data File : VY010753.D
Acq On : 30 Sep 2022 16:38
Operator : KP/MD
Sample : N4898-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CRUSHED-MASONRY-PILE-11RE

Quant Time: Oct 01 06:19:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

