

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
 Data File : VY010754.D
 Acq On : 30 Sep 2022 17:01
 Operator : KP/MD
 Sample : N4897-01RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CRUSHED-MASONRY-PILE-10RE

Quant Time: Oct 01 06:19:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	244135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	406147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	373876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	190318	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	133127	57.011	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.020%
35) Dibromofluoromethane	7.715	113	9401	3.627	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	7.260%#
50) Toluene-d8	10.178	98	455544	47.706	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.420%
62) 4-Bromofluorobenzene	12.477	95	171741	48.210	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
Data File : VY010754.D
Acq On : 30 Sep 2022 17:01
Operator : KP/MD
Sample : N4897-01RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CRUSHED-MASONRY-PILE-10RE

Quant Time: Oct 01 06:19:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

