

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
 Data File : VY010757.D
 Acq On : 30 Sep 2022 18:09
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 01 06:19:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.279	0.190	31.9#	72	0.00
3 P	Chloromethane	0.391	0.283	27.6#	70	0.00
4 C	Vinyl Chloride	0.490	0.379	22.7#	72	0.00
5 T	Bromomethane	0.350	0.275	21.4	77	0.00
6 T	Chloroethane	0.352	0.294	16.5	77	0.00
7 T	Trichlorofluoromethane	0.657	0.581	11.6	81	0.00
8 T	Diethyl Ether	0.225	0.214	4.9	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.395	0.373	5.6	83	0.00
10 T	Methyl Iodide	0.457	0.402	12.0	77	0.00
11 T	Tert butyl alcohol	0.049	0.046	6.1	98	0.00
12 CM	1,1-Dichloroethene	0.353	0.305	13.6#	78	0.00
13 T	Acrolein	0.022	0.019	13.6	80	0.00
14 T	Allyl chloride	0.532	0.507	4.7	85	0.00
15 T	Acrylonitrile	0.117	0.132	-12.8	99	0.00
16 T	Acetone	0.098	0.099	-1.0	82	0.00
17 T	Carbon Disulfide	0.788	0.522	33.8#	66	0.01
18 T	Methyl Acetate	0.316	0.298	5.7	97	0.00
19 T	Methyl tert-butyl Ether	1.101	1.146	-4.1	89	0.00
20 T	Methylene Chloride	0.785	0.457	41.8#	71	0.00
21 T	trans-1,2-Dichloroethene	0.397	0.352	11.3	80	0.00
22 T	Diisopropyl ether	1.257	1.343	-6.8	92	0.00
23 T	Vinyl Acetate	0.742	0.832	-12.1	93	0.00
24 P	1,1-Dichloroethane	0.754	0.772	-2.4	90	0.00
25 T	2-Butanone	0.153	0.166	-8.5	94	0.00
26 T	2,2-Dichloropropane	0.725	0.709	2.2	86	0.00
27 T	cis-1,2-Dichloroethene	0.495	0.487	1.6	86	0.00
28 T	Bromochloromethane	0.301	0.368	-22.3	83	0.00
29 T	Tetrahydrofuran	0.094	0.103	-9.6	97	0.00
30 C	Chloroform	0.816	0.856	-4.9#	91	0.00
31 T	Cyclohexane	0.617	0.484	21.6	77	0.00
32 T	1,1,1-Trichloroethane	0.726	0.729	-0.4	87	0.00
33 S	1,2-Dichloroethane-d4	0.478	0.487	-1.9	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.319	0.299	6.3	83	0.00
36 T	1,1-Dichloropropene	0.356	0.325	8.7	82	0.00
37 T	Ethyl Acetate	0.205	0.223	-8.8	94	0.00
38 T	Carbon Tetrachloride	0.409	0.395	3.4	84	0.00
39 T	Methylcyclohexane	0.416	0.347	16.6	77	0.00
40 TM	Benzene	1.060	1.019	3.9	86	0.00
41 T	Methacrylonitrile	0.105	0.109	-3.8	88	0.00
42 TM	1,2-Dichloroethane	0.326	0.332	-1.8	89	0.00
43 T	Isopropyl Acetate	0.376	0.414	-10.1	95	0.00
44 TM	Trichloroethene	0.303	0.285	5.9	84	0.00
45 C	1,2-Dichloropropane	0.274	0.282	-2.9#	90	0.00
46 T	Dibromomethane	0.161	0.162	-0.6	88	0.00
47 T	Bromodichloromethane	0.398	0.432	-8.5	93	0.00
48 T	Methyl methacrylate	0.166	0.190	-14.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
 Data File : VY010757.D
 Acq On : 30 Sep 2022 18:09
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 01 06:19:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	90	0.00
50 S	Toluene-d8	1.087	1.123	-3.3	78	0.00
51 T	4-Methyl-2-Pentanone	0.213	0.249	-16.9	100	0.00
52 CM	Toluene	0.694	0.673	3.0#	85	0.00
53 T	t-1,3-Dichloropropene	0.402	0.419	-4.2	90	0.00
54 T	cis-1,3-Dichloropropene	0.448	0.456	-1.8	88	0.00
55 T	1,1,2-Trichloroethane	0.237	0.257	-8.4	94	0.00
56 T	Ethyl methacrylate	0.307	0.334	-8.8	92	0.00
57 T	1,3-Dichloropropane	0.392	0.415	-5.9	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.192	-23.9	93	0.00
59 T	2-Hexanone	0.144	0.171	-18.8	99	0.00
60 T	Dibromochloromethane	0.290	0.314	-8.3	92	0.00
61 T	1,2-Dibromoethane	0.222	0.226	-1.8	89	0.00
62 S	4-Bromofluorobenzene	0.474	0.389	17.9	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.334	0.305	8.7	82	0.00
65 PM	Chlorobenzene	0.865	0.853	1.4	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.337	0.356	-5.6	89	0.00
67 C	Ethyl Benzene	1.518	1.509	0.6#	87	0.00
68 T	m/p-Xylenes	0.588	0.579	1.5	87	0.00
69 T	o-Xylene	0.573	0.574	-0.2	87	0.00
70 T	Styrene	0.974	1.009	-3.6	89	0.00
71 P	Bromoform	0.209	0.229	-9.6	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.090	3.030	1.9	88	0.00
74 T	N-amyl acetate	0.749	0.810	-8.1	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.630	0.660	-4.8	96	0.00
76 T	1,2,3-Trichloropropane	0.477	0.475	0.4	97	0.00
77 T	Bromobenzene	0.735	0.707	3.8	88	0.00
78 T	n-propylbenzene	3.681	3.616	1.8	89	0.00
79 T	2-Chlorotoluene	2.097	2.072	1.2	90	0.00
80 T	1,3,5-Trimethylbenzene	2.611	2.562	1.9	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.220	-2.3	94	0.00
82 T	4-Chlorotoluene	2.166	2.163	0.1	91	0.00
83 T	tert-Butylbenzene	2.335	2.364	-1.2	91	0.00
84 T	1,2,4-Trimethylbenzene	2.594	2.580	0.5	90	0.00
85 T	sec-Butylbenzene	3.420	3.460	-1.2	91	0.00
86 T	p-Isopropyltoluene	2.858	2.900	-1.5	90	0.00
87 T	1,3-Dichlorobenzene	1.492	1.477	1.0	90	0.00
88 T	1,4-Dichlorobenzene	1.477	1.461	1.1	90	0.00
89 T	n-Butylbenzene	2.624	2.682	-2.2	91	0.00
90 T	Hexachloroethane	0.549	0.558	-1.6	91	0.00
91 T	1,2-Dichlorobenzene	1.329	1.342	-1.0	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.105	0.113	-7.6	98	0.00
93 T	1,2,4-Trichlorobenzene	0.811	0.802	1.1	85	0.00
94 T	Hexachlorobutadiene	0.493	0.484	1.8	85	0.00
95 T	Naphthalene	1.634	1.739	-6.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
Data File : VY010757.D
Acq On : 30 Sep 2022 18:09
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 01 06:19:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.706	0.715	-1.3	86	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6