

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
 Data File : VY010749.D
 Acq On : 30 Sep 2022 15:08
 Operator : KP/MD
 Sample : N4912-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-211

Quant Time: Oct 01 06:18:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	182999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	289985	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	265230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	137068	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	102719	58.685	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.380%
35) Dibromofluoromethane	7.716	113	98245	53.086	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.180%
50) Toluene-d8	10.179	98	316095	46.451	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.900%
62) 4-Bromofluorobenzene	12.477	95	120474	47.315	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.640%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY093022\
Data File : VY010749.D
Acq On : 30 Sep 2022 15:08
Operator : KP/MD
Sample : N4912-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211

Quant Time: Oct 01 06:18:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY010749.D
Acq: 30 Sep 2022 15:08




Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211


Tgt Ion:168 Resp: 182999
Ion Ratio Lower Upper
168 100
99 52.1 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 58.685 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.001 min
Lab File: VY010749.D
Acq: 30 Sep 2022 15:08

Tgt Ion: 65 Resp: 102719
Ion Ratio Lower Upper
65 100
67 53.5 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY010749.D

Acq: 30 Sep 2022 15:08

Instrument :

MSVOA_Y

ClientSampleId :

VNJ-211





Tgt Ion:114 Resp: 289985

Ion Ratio Lower Upper

114 100

63 19.5 0.0 43.2

88 15.3 0.0 31.8



#35

Dibromofluoromethane

Concen: 53.086 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.006 min

Lab File: VY010749.D

Acq: 30 Sep 2022 15:08




Tgt Ion:113 Resp: 98245


Ion Ratio Lower Upper

113 100

111 102.4 82.6 124.0

192 19.9 17.6 26.4









#50
Toluene-d8
Concen: 46.451 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY010749.D
Acq: 30 Sep 2022 15:08




Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211


Tgt Ion: 98 Resp: 316095
Ion Ratio Lower Upper
98 100
100 65.5 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 47.315 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY010749.D
Acq: 30 Sep 2022 15:08

Tgt Ion: 95 Resp: 120474
Ion Ratio Lower Upper
95 100
174 85.7 0.0 177.0
176 81.6 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY010749.D
Acq: 30 Sep 2022 15:08

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211

Tgt Ion:117 Resp: 265230
Ion Ratio Lower Upper
117 100
82 52.7 43.0 64.4
119 32.5 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY010749.D
Acq: 30 Sep 2022 15:08

Tgt Ion:152 Resp: 137068
Ion Ratio Lower Upper
152 100
115 57.8 28.2 84.7
150 156.5 0.0 347.6

