

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100121\
 Data File : VY006255.D
 Acq On : 01 Oct 2021 11:32
 Operator : SY/MD
 Sample : M3999-01
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 229-342

Quant Time: Oct 02 02:10:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	60810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	99986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	95706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	42723	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	42447	65.218	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.440%
35) Dibromofluoromethane	7.722	113	31802	59.534	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.060%
50) Toluene-d8	10.185	98	125258	56.027	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.060%
62) 4-Bromofluorobenzene	12.483	95	38064	50.118	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100121\
Data File : VY006255.D
Acq On : 01 Oct 2021 11:32
Operator : SY/MD
Sample : M3999-01
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
229-342

Quant Time: Oct 02 02:10:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

