

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100121\
 Data File : VY006262.D
 Acq On : 01 Oct 2021 15:43
 Operator : SY/MD
 Sample : M4035-03
 Misc : 5.69G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 286

Quant Time: Oct 02 02:13:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	124302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	200353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	185505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85995	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	79539	59.786	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.580%
35) Dibromofluoromethane	7.728	113	59223	55.328	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.660%
50) Toluene-d8	10.185	98	250342	55.882	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.760%
62) 4-Bromofluorobenzene	12.483	95	78883	51.833	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100121\
Data File : VY006262.D
Acq On : 01 Oct 2021 15:43
Operator : SY/MD
Sample : M4035-03
Misc : 5.69G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
286

Quant Time: Oct 02 02:13:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

