

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
 Data File : VY015775.D
 Acq On : 02 Oct 2023 13:40
 Operator : SY/MD
 Sample : VY1002SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1002SBL01

Quant Time: Oct 03 01:20:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	189092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	320783	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	288014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125629	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	90681	43.585	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.180%
35) Dibromofluoromethane	7.716	113	85059	44.302	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.600%
50) Toluene-d8	10.179	98	317145	48.716	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.440%
62) 4-Bromofluorobenzene	12.483	95	101166	39.599	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
Data File : VY015775.D
Acq On : 02 Oct 2023 13:40
Operator : SY/MD
Sample : VY1002SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1002SBL01

Quant Time: Oct 03 01:20:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

