

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
 Data File : VY015778.D
 Acq On : 02 Oct 2023 15:41
 Operator : SY/MD
 Sample : 04642-01
 Misc : 4.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PAT-01-09272023

Quant Time: Oct 02 16:59:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	143995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	251038	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	229487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	90442	57.085	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.160%
35) Dibromofluoromethane	7.710	113	80080	53.296	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.600%
50) Toluene-d8	10.179	98	293693	57.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.300%
62) 4-Bromofluorobenzene	12.483	95	96768	48.401	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.800%
Target Compounds						
20) Methylene Chloride	4.710	84	20845	7.070	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
Data File : VY015778.D
Acq On : 02 Oct 2023 15:41
Operator : SY/MD
Sample : 04642-01
Misc : 4.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PAT-01-09272023

Quant Time: Oct 02 16:59:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

