

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
 Data File : VY015780.D
 Acq On : 02 Oct 2023 16:30
 Operator : SY/MD
 Sample : 04691-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-08-092923

Quant Time: Oct 03 01:23:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	134609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	233758	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	213104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	94130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	82946	56.004	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.000%
35) Dibromofluoromethane	7.716	113	75902	54.250	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.500%
50) Toluene-d8	10.179	98	275051	57.980	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.960%
62) 4-Bromofluorobenzene	12.483	95	89684	48.174	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.340%
Target Compounds						
20) Methylene Chloride	4.710	84	19673	7.190	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
Data File : VY015780.D
Acq On : 02 Oct 2023 16:30
Operator : SY/MD
Sample : 04691-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-08-092923

Quant Time: Oct 03 01:23:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

