

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
 Data File : VY015782.D
 Acq On : 02 Oct 2023 17:27
 Operator : SY/MD
 Sample : 04627-27
 Misc : 1.58g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 280-300FB-BACKFILL-14

Quant Time: Oct 03 01:24:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	153950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	277616	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	254995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	115858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	82120	48.480	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.960%
35) Dibromofluoromethane	7.716	113	73960	44.511	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.020%
50) Toluene-d8	10.179	98	267966	47.563	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.120%
62) 4-Bromofluorobenzene	12.477	95	89838	40.633	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.260%
Target Compounds						
20) Methylene Chloride	4.710	84	25019	8.595	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
Data File : VY015782.D
Acq On : 02 Oct 2023 17:27
Operator : SY/MD
Sample : 04627-27
Misc : 1.58g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
280-300FB-BACKFILL-14

Quant Time: Oct 03 01:24:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

