

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
 Data File : VY015785.D
 Acq On : 02 Oct 2023 18:38
 Operator : SY/MD
 Sample : 04657-02RE
 Misc : 5.90g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 U-2(2-6IN)RE

Quant Time: Oct 03 01:25:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	128290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	223644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	199595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	76845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	82551	58.483	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.960%
35) Dibromofluoromethane	7.716	113	73121	54.626	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.260%
50) Toluene-d8	10.179	98	263188	57.988	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.980%
62) 4-Bromofluorobenzene	12.483	95	78858	44.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.540%
Target Compounds						
95) Naphthalene	15.233	128	14943	3.513	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
Data File : VY015785.D
Acq On : 02 Oct 2023 18:38
Operator : SY/MD
Sample : 04657-02RE
Misc : 5.90g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
U-2(2-6IN)RE

Quant Time: Oct 03 01:25:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

