

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
 Data File : VY015786.D
 Acq On : 02 Oct 2023 19:01
 Operator : SY/MD
 Sample : 04657-04
 Misc : 5.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 U-1(2-6IN)

Quant Time: Oct 03 01:26:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	127585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	224141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	199061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	77241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	80690	57.480	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.960%
35) Dibromofluoromethane	7.716	113	73303	54.640	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.280%
50) Toluene-d8	10.179	98	264221	58.087	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.180%
62) 4-Bromofluorobenzene	12.483	95	79937	44.780	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.560%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	18890	7.353	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100223\
Data File : VY015786.D
Acq On : 02 Oct 2023 19:01
Operator : SY/MD
Sample : 04657-04
Misc : 5.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
U-1(2-6IN)

Quant Time: Oct 03 01:26:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

