

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
 Data File : VY015801.D
 Acq On : 03 Oct 2023 13:22
 Operator : SY/MD
 Sample : 04627-27
 Misc : 3.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 280-300FB-BACKFILL-14

Quant Time: Oct 04 01:10:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	110123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	191932	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	176972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	81326	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	76205	62.893	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.780%
35) Dibromofluoromethane	7.716	113	63899	55.623	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.240%
50) Toluene-d8	10.179	98	225996	58.021	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.040%
62) 4-Bromofluorobenzene	12.483	95	76114	49.794	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
Data File : VY015801.D
Acq On : 03 Oct 2023 13:22
Operator : SY/MD
Sample : 04627-27
Misc : 3.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
280-300FB-BACKFILL-14

Quant Time: Oct 04 01:10:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

