

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
 Data File : VY015804.D
 Acq On : 03 Oct 2023 14:33
 Operator : SY/MD
 Sample : 04685-02
 Misc : 5.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PP-20230928-B4A-0-2

Quant Time: Oct 04 01:11:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	119332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	211046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	192389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	81283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	78423	59.729	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.460%
35) Dibromofluoromethane	7.716	113	69141	54.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.480%
50) Toluene-d8	10.179	98	247539	57.796	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.600%
62) 4-Bromofluorobenzene	12.483	95	80250	47.745	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.500%
Target Compounds						
20) Methylene Chloride	4.704	84	39734	23.230	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
Data File : VY015804.D
Acq On : 03 Oct 2023 14:33
Operator : SY/MD
Sample : 04685-02
Misc : 5.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PP-20230928-B4A-0-2

Quant Time: Oct 04 01:11:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

