

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
 Data File : VY015807.D
 Acq On : 03 Oct 2023 15:44
 Operator : SY/MD
 Sample : 04685-08
 Misc : 6.78g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PP-20230928-B6A-13-15

Quant Time: Oct 04 01:12:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	190162	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	187935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	91986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	75530	65.881	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 131.760%	
35) Dibromofluoromethane	7.710	113	63105	55.444	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.880%	
50) Toluene-d8	10.179	98	224598	58.198	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 116.400%	
62) 4-Bromofluorobenzene	12.477	95	84011	55.472	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 110.940%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	27996	17.710	ug/l	90
95) Naphthalene	15.233	128	9875	1.939	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
Data File : VY015807.D
Acq On : 03 Oct 2023 15:44
Operator : SY/MD
Sample : 04685-08
Misc : 6.78g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PP-20230928-B6A-13-15

Quant Time: Oct 04 01:12:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

