

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
 Data File : VY015812.D
 Acq On : 03 Oct 2023 17:42
 Operator : SY/MD
 Sample : 04703-01
 Misc : 5.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 E9066-OB-001

Quant Time: Oct 04 01:13:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	165452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	281748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	236582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	82802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	99119	54.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.900%	
35) Dibromofluoromethane	7.716	113	85712	50.827	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.660%	
50) Toluene-d8	10.179	98	331608	57.995	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 116.000%	
62) 4-Bromofluorobenzene	12.477	95	92834	41.372	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 82.740%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.192	76	11465	4.054	ug/l	97
20) Methylene Chloride	4.704	84	33059	11.797	ug/l	89
40) Benzene	8.155	78	9158	1.204	ug/l	95
52) Toluene	10.240	92	8217	1.676	ug/l	95
64) Tetrachloroethene	10.715	164	10952	5.396	ug/l	91
68) m/p-Xylenes	11.703	106	6135	1.788	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100323\
Data File : VY015812.D
Acq On : 03 Oct 2023 17:42
Operator : SY/MD
Sample : 04703-01
Misc : 5.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
E9066-OB-001

Quant Time: Oct 04 01:13:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

