

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
 Data File : VY019782.D
 Acq On : 03 Oct 2024 12:57
 Operator : SY/MD
 Sample : P4277-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-03-10022024

Quant Time: Oct 04 01:25:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	5012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	9143	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	6066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	3459	74.531	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.060%
35) Dibromofluoromethane	7.634	113	2976	57.408	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.820%
50) Toluene-d8	10.109	98	9997	52.138	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.280%
62) 4-Bromofluorobenzene	12.408	95	1302	17.660	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	35.320%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
Data File : VY019782.D
Acq On : 03 Oct 2024 12:57
Operator : SY/MD
Sample : P4277-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-03-10022024

Quant Time: Oct 04 01:25:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

