

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
 Data File : VY019783.D
 Acq On : 03 Oct 2024 13:21
 Operator : SY/MD
 Sample : P4277-02
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-03-10022024

Quant Time: Oct 04 01:26:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	5538	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	10243	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	6221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	4125	80.440	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	160.880%
35) Dibromofluoromethane	7.634	113	3225	55.531	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.060%
50) Toluene-d8	10.109	98	10956	51.003	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.402	95	1532	18.548	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	37.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
Data File : VY019783.D
Acq On : 03 Oct 2024 13:21
Operator : SY/MD
Sample : P4277-02
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-03-10022024

Quant Time: Oct 04 01:26:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

