

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
 Data File : VY019785.D
 Acq On : 03 Oct 2024 14:07
 Operator : SY/MD
 Sample : P4273-02
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-05-100224

Quant Time: Oct 04 01:27:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	39993	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	66530	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	46000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.341	152	11609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	17449	47.118	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.240%
35) Dibromofluoromethane	7.634	113	19610	51.987	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.980%
50) Toluene-d8	10.103	98	72338	51.847	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.700%
62) 4-Bromofluorobenzene	12.402	95	16819	31.351	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	62.700%
Target Compounds						
16) Acetone	3.867	43	10037	113.766	ug/l	# Qvalue 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100324\
Data File : VY019785.D
Acq On : 03 Oct 2024 14:07
Operator : SY/MD
Sample : P4273-02
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-05-100224

Quant Time: Oct 04 01:27:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

